


Reconstituting the IL-12 (p70) standard in 2 mL of assay diluent results in a concentration of 5000 pg/mL. Standard curves for no-wash and 1-wash protocols are shown starting from 5000 pg/mL and proceeding in a 1:2 serial titration. The limit of detection, represented by the dashed line, is determined by adding 3 standard deviations to the mean of samples without IL-12 (p70) (N=12). Results are typical and may vary based on specific assay conditions.

No-Wash Assay Performance





Lowest Detectable Concentration Tested (pg/mL): 20

For research purposes only. Not for resale.

Lowest Detectable Concentration Tested (pg/mL): 10

Rev. 1 Sep 2014